

Atlantis® FLO compatibility chart

The Atlantis FLO (Feature Locating Object) is used as a digital transfer coping, ensuring that the position and orientation of the implant is accurately represented in the Atlantis VAD (Virtual Abutment Design) software. The FLO is used together with the scanners enabling implant cases to be scanned in the dental laboratory and digital files to be sent to Dentsply Sirona Implants for production of Atlantis abutments.

For simplified handling, the FLO screw is designed to be “captured” in the FLO to prevent it from falling out.



Atlantis FLO Kit Select


Includes FLOs for the most frequently scanned interfaces.

Atlantis FLO series

Includes all FLOs within each of the interface groups categorized by the same letter.

Atlantis FLO singles

Available for purchase as a single item, unless otherwise noted.



Product	REF	FLO code
Atlantis FLO Kit Select	34603	See FLOs included in the table on next page
Atlantis FLO A-series	34604	A-03, A-04, A-06, A-07, A-09, A-11
Atlantis FLO B-series	34605	B-01, B-02, B-03, B-04
Atlantis FLO C-series	34606	C-01, C-02, C-03, C-04
Atlantis FLO D-series	34607	D-01, D-02, D-03, D-04
Atlantis FLO E-series	34608	E-01, E-02
Atlantis FLO F-series	34609	F-01, F-02
Atlantis FLO H-series	34610	H-01, H-02
Atlantis FLO J-series	34611	J-01, J-02, J-03
Atlantis FLO K-series	34612	K-01, K-02, K-03, K-04
Atlantis FLO M-series	34613	M-01, M-02, M-03
Atlantis FLO N-series	34614	N-01, N-02, N-03, N-04
Atlantis FLO P-series	34660	P-01, P-02, P-03, P-04, P-05, P-06, P-07
Atlantis FLO Q-series	34691	Q-01, Q-02, Q-03
Atlantis FLO R-series	34692	R-01, R-02
Atlantis FLO singles	-	FLOs available as single items are listed in the table on next page
Atlantis FLO case	34602	-
Atlantis FLO screwdriver	34601	-

Manufacturer	Implant	Atlantis FLO Kit Select	FLO code	REF
BioHorizons	Internal, Tapered Internal 3.5	x	B-01	34527
	Internal, Tapered Internal 4.5	x	B-02	34528
	Internal, Tapered Internal 5.7	x	B-03	34529
	External Implant 3.5		A-03	34518
	External Implant 4.0		A-06	34521
	External Implant 5.0		A-09	34524
Biomet 3i	Certain 3.25, 4/3	x	F-01	34539
	Certain 4.0, 5.0, 6.0, 5/4	x	F-02	34540
	Certain Prevail 3/4/3 (3.4), 4/3	x	F-01	34539
	Certain Prevail 4/5/4, 5/4, 5/6/5, 6/5	x	F-02	34540
	Certain XP 4/5, 5/6	x	F-02	34540
	Miniplant 3.25		A-07	34522
	Osseotite 3.75, 4.0		A-07	34522
	Osseotite 5.0, 6.0		A-07	34522
Camlog	Osseotite XP 4/5, 5/6		A-07	34522
	CONELOG 3.3		Q-01	34666
	CONELOG 3.8, 4.3		Q-02	34667
	CONELOG 5.0		Q-03	34668
	Screw-line Implant 3.3		N-01	34584
	Screw-line/Root-line Implant 3.8		N-02	34585
	Screw-line/Root-line Implant 4.3		N-03	34586
	Screw-line/Root-line Implant 5.0		N-04	34587
Dentsply Sirona	Screw-line/Root-line Implant 6.0		N-04	34587
	Ankylos C/X - C/ 3.5, 4.5, 5.5, 7.0 (non-indexed)	x	L-01	34583
	Ankylos C/X - /X 3.5, 4.5, 5.5, 7.0 (indexed)	x	L-01	34583
	OsseoSpeed EV 3.0 - Green	x	P-01	34593
	OsseoSpeed EV 3.6 - Purple	x	P-02	34594
	OsseoSpeed EV 4.2 - Yellow	x	P-03	34595
	OsseoSpeed EV 4.8 - Blue	x	P-04	34596
	OsseoSpeed EV 5.4 - Brown	x	P-05	34597
	OsseoSpeed Profile EV 4.2 Yellow	x	P-06	34598
	OsseoSpeed Profile EV 4.8 Blue	x	P-07	34599
	OsseoSpeed/OsseoSpeed TX 3.0 - Yellow*	x	C-01	34530
	OsseoSpeed/OsseoSpeed TX 3.5, 4.0 - Aqua*	x	C-02	34531
	OsseoSpeed/OsseoSpeed TX 4.5, 5.0 - Lilac*	x	C-03	34532
	OsseoSpeed TX Profile 4.5, 5.0*	x	C-04	34592
	Xive S 3.0	x	J-01	34575
	Xive S 3.4	x	J-02	34576
	Xive S 3.8, 4.5, 5.5	x	J-03	34577
	Frialit 3.4	x	J-02	34576
	Frialit 3.8, 4.5, 5.5	x	J-03	34577
	Keystone Dental	PrimaConnex SD 3.3, Genesis 3.8		M-01
PrimaConnex SD 3.5			M-01	34589
PrimaConnex RD 4.0, Genesis 4.5			M-02	34590
PrimaConnex RD 4.1			M-02	34590
PrimaConnex WD 5.0, Genesis 5.5, 6.5			M-03	34591
MIS	M4 NP 3.3		B-04	34678
	M4 SP 3.75, 4.2		B-01	34527
	M4 WP 5.0, 6.0		B-02	34528
	SEVEN NP 3.3		B-04	34678
	SEVEN SP 3.75, 4.2		B-01	34527
	SEVEN WP 5.0, 6.0		B-02	34528

All trademarks and company names are the property of their respective owners.

*A light coat of petroleum jelly on the tapered surface of these FLOs is recommended for ease of removal after scanning.

Manufacturer	Implant	Atlantis FLO Kit Select	FLO code	REF
Nobel Biocare	NobelActive 3.0		K-03	34665
	NobelActive/NobelReplace Conical Connection NP 3.5		K-01	34578
	NobelActive/NobelReplace Conical Connection RP 4.3, 5.0		K-02	34579
	NobelActive WP 5.5		K-04	34677
	NobelReplace NP 3.5	x	D-01	34533
	NobelReplace RP 4.3	x	D-02	34534
	NobelReplace WP 5.0	x	D-03	34535
	NobelReplace 6.0	x	D-04	34536
	Brånemark System NP 3.3		A-04	34519
	Brånemark System RP 3.75, 4.0		A-06	34521
	Brånemark System WP 5.0, 5.5		A-11	34526
Osstem	Osstem TS Mini 3.5		R-01	34671
	Osstem TS Regular 4.0, 4.5, 5.0, 6.0, 7.0		R-02	34672
Straumann	Bone Level 3.3 NC (Narrow CrossFit)*	x	H-01	34542
	Bone Level 4.1, 4.8 RC (Regular CrossFit)*	x	H-02	34543
	Standard/Standard Plus 4.8 RN (Regular Neck)		E-01	34537
	Standard/Standard Plus 4.8 WN (Wide Neck)		E-02	34580
	Standard Plus 3.3 NN (Narrow Neck)		G-01	34541
Zimmer Dental	Tapered Screw-Vent 3.5/Tapered Screw-Vent 3.5 friction fit	x	B-01	34527
	Tapered Screw-Vent 4.5/Tapered Screw-Vent 4.5 friction fit	x	B-02	34528
	Tapered Screw-Vent 5.7/Tapered Screw-Vent 5.7 friction fit	x	B-03	34529
	Screw-Vent 3.3, 3.7/Screw-Vent 3.3, 3.7 friction fit	x	B-01	34527
	Screw-Vent 4.5/Screw-Vent 4.5 friction fit	x	B-02	34528
	SwissPluss 4.8		E-01	34537
Atlantis FLO case	—	x	—	34602
Atlantis FLO screwdriver	—	x	—	34601
Atlantis FLO tag set	—		—	34664

All trademarks and company names are the property of their respective owners.

*A light coat of petroleum jelly on the tapered surface of these FLOs is recommended for ease of removal after scanning.